



NON-VOLATILE ACIDS (Sulfuric Acid and Phosphoric Acid) 7908

Formulae: H ₂ SO ₄	MW: 98.08	CAS: 7664-93-9	RTECS: WS5600000
H ₃ PO ₄	MW: 98.00	CAS: 7664-38-2	RTECS: TB6300000

METHOD: 7908, Issue 2

EVALUATION: FULL

Issue 1: 10 May 2014

Issue 2: 16 November 2023

OSHA: 1 mg/m³ (H₂SO₄ & H₃PO₄)
NIOSH: 1 mg/m³ (H₂SO₄ & H₃PO₄); STEL: 3 mg/m³ (H₃PO₄)
OTHER OELS: [1,2]

PROPERTIES [3]:

H₂SO₄: Liquid, mp = 10 °C; bp = 337°C
H₃PO₄: Solid (often used in an aqueous solution), mp = 42.4°C; bp = 407 °C

SYNONYMS: H₂SO₄: hydrogen sulfate, oil of vitriol; H₃PO₄: ortho-phosphoric acid

SAMPLING	MEASUREMENT
<p>SAMPLER: FILTER, 37-mm diameter quartz fiber; or polytetrafluoroethylene (PTFE), 0.45 µm pore size</p> <p>FLOW RATE: 1 to 5 L/min</p> <p>VOL-MIN: 15 L -MAX: 1000 L</p> <p>SHIPMENT: 4 °C</p> <p>SAMPLE STABILITY: Filter: at least 7 days at room temp. (See Evaluation section); Extract: 1 week at room temp. and at 4 °C thereafter to 28 days [4,5]</p> <p>BLANKS: 3 field blanks minimum per set</p>	<p>TECHNIQUE: ION CHROMATOGRAPHY (IC) with conductivity detection</p> <p>ANALYTES: Sulfate (SO₄²⁻) ion, phosphate (PO₄³⁻) ion</p> <p>EXTRACTION: Aqueous solution of sodium carbonate / sodium hydrogen carbonate</p> <p>INJECTION VOLUME: 50 µL</p> <p>ELUENT: 2.7 mM Na₂CO₃/0.3 mM NaHCO₃, flow rate 1.5 mL/min</p> <p>COLUMNS: Pre-column, anion-exchange column and suppressor column</p> <p>CALIBRATION RANGE: H₂SO₄ = 0.06 mg/L to 10 mg/L; H₃PO₄ = 0.1 mg/L to 10 mg/L</p> <p>ESTIMATED LOD: H₂SO₄ = 0.001 mg/sample (quartz), 0.0003 mg/sample (PTFE); H₃PO₄ = 0.0005 mg/sample (quartz or PTFE) [8]</p> <p>PRECISION (\bar{S}_T): H₂SO₄ = 0.043; H₃PO₄ = 0.032 [4,6]</p>
ACCURACY*	
<p>RANGE STUDIED: 0.005 to 2.0 mg/sample</p> <p>BIAS*: Negligible [5]</p> <p>OVERALL PRECISION (\bar{S}_{TT}): H₂SO₄ = 0.107; H₃PO₄ = 0.106 [6]</p> <p>EXPANDED UNCERTAINTY*: less than 23% for both H₂SO₄ and H₃PO₄ [5]</p> <p>* Accuracy calculations were determined using references 5 and 7 rather than the traditional NIOSH accuracy criteria.</p>	

APPLICABILITY: The working range for H₂SO₄ is (at least) 0.008 to 1.0 mg/sample and for H₃PO₄ is 0.004 to 1.0 mg/sample for a 420 Liter air sample [6,8]. This method has been updated to include sample storage study and filter variability information.

INTERFERENCES: Particulate salts of sulfate or phosphate will give a positive interference. Both quartz fiber and PTFE filters may have a variable sulfate background.

OTHER METHODS: This procedure, which is consistent with ISO 21438-1 [5], replaces NIOSH 7903 [9] for the determination of sulfuric and phosphoric acid in workplace air samples by IC. The main advantage of NIOSH method 7908 is that it can allow for the collection of the inhalable fraction of sulfuric and phosphoric acid aerosols by means of the pre-filter (housed within an optional inhalable sampler).

REAGENTS:

1. Water, deionized (DI), ≥ 18 M Ω -cm resistivity
2. Sodium carbonate (Na₂CO₃), anhydrous, American Chemical Society (ACS) analytical grade
3. Sodium hydrogen carbonate (NaHCO₃), ACS analytical grade
4. Extraction & eluent stock solution: 0.27 M Na₂CO₃ / 0.03 M NaHCO₃; dissolve 2.86 g Na₂CO₃ and 0.25 g NaHCO₃ in 25 mL of deionized water and swirl to mix. Then bring to 100 mL in a volumetric flask, stopper and mix thoroughly.
5. Extraction & eluent solution: 0.0027 M Na₂CO₃ / 0.0003 M NaHCO₃; transfer 10 mL of 0.27 M Na₂CO₃ / 0.03 M NaHCO₃ stock solution to a 1 L volumetric flask, dilute to the mark with deionized water, stopper and mix thoroughly.
6. Sulfate (SO₄²⁻) ion and phosphate (PO₄³⁻) ion standard solutions, each 1000 mg/L
7. Calibration stock solution, 100 mg/L (as the anion): Place 10 mL aliquots of sulfate and phosphate standard solution into a 100 mL volumetric flask, dilute to the mark and mix thoroughly.

*See SPECIAL PRECAUTIONS

EQUIPMENT:

1. Sampler: filter, 37-mm diameter quartz fiber; or polytetrafluoroethylene (PTFE), 0.45- μ m pore size, in cassette filter holder manufactured from acid-resistant (chemically inert) material
NOTE: Quartz fiber filters should be binderless and heat-treated.
2. Personal sampling pump, 1 to 5 L/min, with flexible connecting tubing
3. Ion chromatograph, with pre-column (50 mm by 4.0 mm), anion-exchange column (200 mm by 4.0 mm), suppressor column (4 mm) and conductivity detector; and connected to data processing unit
4. Ultrasonic bath
5. Vessels, 10-mL or 15-mL, plastic (e.g., polypropylene), with screw caps
6. Volumetric flasks, 25- to 1000-mL
7. Pipets, 50- to 10,000- μ L
8. Beakers, 25- to 100-mL
9. Water purification system, to prepare greater than or equal to 18 M Ω -cm resistivity deionized water
10. Bottles, polyethylene, 100-mL
11. Syringes, plastic, 5-mL
12. Syringe filter cartridges, with 0.8- μ m pore size PTFE membrane filters
13. Micro-syringes, 50- μ L, with 60-mm x 0.6-mm needles
14. Tweezers, PTFE-coated
15. Auto-sampler vials, 2-mL capacity
16. Analytical balance, with capability of weighing to nearest 0.01 mg

SPECIAL PRECAUTIONS: Wear gloves, lab coat, and safety glasses while handling acids. All work should be performed with adequate ventilation for personnel and equipment. It is imperative that acid be added to water in order to avoid a violent exothermic reaction. There is risk of fire and explosion when phosphoric acid has contact with bases, combustible substances, oxidizing agents, reducing agents or water. Avoid physical contact since both sulfuric and phosphoric acid are corrosive and irritants to eyes, skin, and the respiratory system. These are caustic materials and can react with metals to form flammable hydrogen gas. Do NOT mix with solutions containing bleach or ammonia [10,11].

SAMPLING, SAMPLE TRANSPORT AND STORAGE:

1. Calibrate each personal sampling pump with a representative sampler in-line.
2. Sample at an accurately known flow rate between 1 and 5 L/min for a total sample size of 15 to 1000 liters. Avoid sampler overloading.
3. Immediately after sampling, remove the filter from the cassette with PTFE-coated tweezers and place it in a 10-mL screw-cap plastic vessel. With about 2 mL extraction solution (0.0027 M Na₂CO₃ / 0.0003 M NaHCO₃), rinse material from the inside surfaces of the cassette into the vessel. Add extraction solution into the vessel until a final volume of 5 mL is reached.

NOTE: If field desorption is not possible, filters may be stored within the cassettes or transferred to 15-mL polypropylene tubes for shipment to laboratory. However, samples should be prepared in a timely manner after receipt by the laboratory. See Evaluation Section for more information.

4. Submit at least three field blanks for each set of samples collected per day. Handle these in the same way as the field samples; i.e., place each filter into a vessel, add 5 mL of eluent solution and ship it to the lab along with the remaining samples.
5. Refrigerate all samples that are to be stored overnight (or longer) prior to shipment to the laboratory. Ship all samples to the laboratory in accordance with established chain-of-custody procedures [12].
6. Refrigerate the samples (4 °C) immediately upon receipt at the lab until ready for analysis.
7. Analyze samples within 4 weeks of receipt.

SAMPLE PREPARATION:

NOTE: If samples were not field desorbed, begin sample preparation with Step 3.

8. Remove sample vessels from storage and bring them to room temperature.
9. Sonicate the samples in an ultrasonic bath for at least 15 minutes and allow to cool for at least 30 minutes.
10. Using 5-mL syringes, filter each sample extract solution through a PTFE filter into clean plastic vessels or into autosampler vials (if autosampler used).

CALIBRATION AND QUALITY CONTROL:

11. Through dilution of the calibration stock solution in eluent solution, prepare calibration working standard solutions covering the range of approximately 0.06 to 10 mg/L of sulfate and phosphate. Modify the range of calibration standards depending upon the working range for your instrumentation. Store working standards in tightly sealed polyethylene bottles. Prepare fresh working standards weekly.
12. Calibrate the ion chromatograph with at least six working standards covering the range of 0.06 to 10 mg/L of sulfate and phosphate ion per sample by preparing a calibration graph of anion peak height (μS [micro siemens]) vs. concentration (mg/L).
13. Analyze at least one calibration standard per ten samples to confirm that instrument calibration is maintained over the course of the analytical run.
14. Check recoveries with QC samples of at least one media blank (minimum of 3) and at least one spiked media blank per twenty samples (or batch). QC values should be within established control limits. If not, the batch is considered out of control and corrective actions taken before more samples are analyzed.
15. Analyze working standards together with samples, reagent blanks, field blanks, and QC samples.

MEASUREMENT:

16. Set the ion chromatograph to desired eluent flow rate, e.g., 1.5 mL/min, and column pressure, e.g., 1.1×10^5 kPa, and other conditions as specified by the instrument manufacturer.
17. Inject a sample aliquot, e.g., 50- μL , into the chromatograph, and measure the peak heights of the phosphate and sulfate peaks (at retention times of about 9 min and 11.6 min, respectively). If the peak height exceeds the linear calibration range, dilute with eluent, reanalyze, and apply the appropriate dilution factor in calculations.

CALCULATIONS:

18. Calculate the mass concentration of each anion, C (mg/m^3), in the air volume sampled, V (L):

$$C = \left(\frac{((C_1 * V_1 * F_d) - (C_0 * V_0)) * 1000}{V} \right) * F_c$$

where:

C_0 = mean concentration, in mg/L, of anion in the field blank test solutions;

C_1 = concentration, in mg/L, of anion in the sample test solution;

V = volume, in liters, of the air sample;
 V_0 = volume, in L, of the field blank test solutions;
 V_1 = volume, in L, of the sample test solutions;
 F_d = dilution factor for each sample test solution;
 F_c = conversion factor to convert from anion to acid concentration: $F_c = 1.021$ for sulfate; $F_c = 1.031$ for phosphate.

EVALUATION OF METHOD:

Reported recoveries of sulfuric acid were greater than 95% over the range of 0.1 to 2 times the limit value on 0.45- μm pore size PTFE filters [13]. Greater than 95% recovery of extracts of sulfuric acid and phosphoric acid collected on quartz fiber filters was found four weeks after sample collection with extracted samples stored at room temperature for 7 days and under refrigeration thereafter [4]. The validated measuring range extends to 1 mg/m^3 [6]. Mean analytical recovery determined from the analysis of spiked quartz fiber filters has been found to be in the range of 97 to 100% for both acids [4]. The component of the coefficient of variation of the method that arises from analytical variability, determined from the analysis of spiked quartz fiber filters, was 0.7% to 3.2% for phosphoric acid and 0.5% to 2.6% for sulfuric acid [4]. An interlaboratory study with 26 participants found negligible biases and interlaboratory relative standard deviations of 12 to 15% for sulfuric acid and phosphoric acid concentrations between 0.05 and 1 mg/m^3 [14]. The method has also been field tested for sulfuric acid measurements at sample volumes of up to nearly 2,000 liters [15]. The analytical figures of merit for the method satisfy performance criteria specified in an applicable consensus standard [16]. Issue 1 validation data are found in references 4, 6, and 13 and the user check report is found in reference 14.

Issue 2 Revisions:

In this issue, the calibration range is updated to match what was used for the new studies, and information on the storage stability of samples, i.e., the need for field desorption, and filter variability is added.

In ISO 21438-1 [5] samples are reported to be stable for 4 weeks; however, sample preparation includes field desorption of quartz fiber filters and users are cautioned about anecdotal evidence relating to the storage of quartz fiber filters. NIOSH conducted a series of storage stability studies to test the necessity of field desorption, fully described in [8]. In one study quartz fiber filters were spiked with sulfuric and phosphoric acids at 2 levels (10 μg and 40 μg). Filters were stored at ambient temperature, either within a closed-faced cassette (CFC) assembly or in 15-mL polypropylene tubes, and analyzed at Days 0, 7, and 14. For sulfuric acid spikes, sample stability was affected by spike level. For this reason, quartz fiber filter samples containing sulfuric acid should be prepared within 7 days of sampling. Phosphoric acid quantitatively recovered regardless of concentration or storage vessel over the 14 days of the study, however, the recovery did decline over the storage period.

In another study PTFE filters were spiked with sulfuric and phosphoric acid at 4 levels within the analytical range of the method (10, 40, 160, and 1000 μg), separately and in combination [16]. Samples were stored within acid-resistant cassettes at ambient temperature. No significant difference was found between acids spiked separately or in combination. Sample storage stability varied based upon spike level and the results at the 160 and 1000 μg levels were found to be significantly different ($p=0.0060$). For nearly all of the spike levels, the samples were found to be stable for 28 days. Outliers in the reported storage stability study may be due to issues with spiking relatively large acid volumes onto filters and handling spiked filters where the liquid was still present, neither of which would be conditions encountered with field samples.

NIOSH also conducted studies on the background levels of sulfuric and phosphoric acid in quartz fiber and PTFE filters [16]. Quartz fiber filters were tested from three manufacturers. Multiple lots were tested, when available. A sulfuric acid background was found on all filters, with % RSDs ranging from 7 to 69%. No phosphoric acid was detected on the quartz fiber filters. PTFE filters were tested from two manufacturers. Multiple lots were tested, when available. One lot (of three tested) from a single manufacturer had a variable sulfuric acid background (110% RSD) and one filter from the other manufacturer had a sulfuric acid response.

In each instance, the responses were near the LOD. No phosphoric acid background was detected on the PTFE filters. Method users are encouraged to adequately assess and understand the background response when analyzing samples for sulfuric acid.

REFERENCES:

- [1] ACGIH [2023]. TLVs and BEIs based on the documentation of the Threshold Limit Values for chemical substances and physical agents and Biological Exposure Indices. Cincinnati, Ohio: American Conference of Governmental Industrial Hygienists. [www.acgih.org]
- [2] Institute for Occupational Safety and Health of the German Social Accident Insurance (IFA) [no date]. GESTIS international limit values, <https://limitvalue.ifa.dguv.de/>. Date accessed: November 16, 2023.
- [3] Haynes, WH, ed. [2013]. CRC Handbook of chemistry and physics. 94th ed. Boca Raton, FL: CRC Press.
- [4] Breuer D [2002]. Inorganic acid mists (H₂SO₄, H₃PO₄). In: Kettrup A, Greim H, eds. Analyses of hazardous substances in air (Deutsche Forschungsgemeinschaft). Vol. 6. Weinheim, FRG: Wiley-VCH. pp. 67-78.
- [5] ISO [2021]. ISO 21438-1: Workplace atmospheres – Determination of inorganic ions by ion chromatography – Part 1: Non-volatile acids (sulfuric acid and phosphoric acid). Geneva, Switzerland: International Organization for Standardization (ISO).
- [6] Breuer D [2013]. Anorganische säuren, partikulär: phosphorsäure, schwefelsäure (Inorganic acid mists: phosphoric acid, sulfuric acid), in IFA Working Folder, Sheet No. 6173. Berlin, FRG: Erich- Schmidt Verlag. ISBN 978-3-503-13084-9.
- [7] ISO (International Organization for Standardization) [1995]. ISO Guide 98: Guide to the expression of uncertainty in measurement. Geneva, Switzerland: ISO.
- [8] NIOSH [2023]. NIOSH 7908 Backup Data Report. In: Andrews, RN, ed. NIOSH Manual of analytical methods, 5th ed., Cincinnati, OH: U.S. Department of Health and Human Services, Centers for Disease Control and Prevention, National Institute for Occupational Safety and Health, DHHS (NIOSH). Publication No. 2014-151. [www.cdc.gov/niosh/nmam/datadoc.html]
- [9] NIOSH [1994]. Acids, inorganic: Method 7903. In: Eller PM, Cassinelli ME, eds. NIOSH Manual of analytical methods, 4th ed., Cincinnati, OH: U.S. Department of Health and Human Services, Centers for Disease Control and Prevention, National Institute for Occupational Safety and Health, DHHS (NIOSH). Publication No. 94-113. [www.cdc.gov/niosh/nmam]
- [10] NIOSH [2007]. NIOSH Pocket guide to chemical hazards. Barsan ME, ed. Cincinnati, OH: US Department of Health and Human Services, Centers for Disease Control and Prevention, National Institute for Occupational Safety and Health, DHHS (NIOSH) Publication No. 2005-149.
- [11] Furr AK, ed. [1995] CRC Handbook of laboratory safety, 4th ed. Boca Raton, FL: CRC Press.
- [12] ASTM International [2010]. ASTM D4840: Standard guide for sampling chain of custody procedures. West Conshohocken, PA: ASTM International.
- [13] Breuer D [1999]. Measurement of vapor-aerosol mixtures. *J Environ Monit* 1:299-305.
- [14] Breuer D, Howe A [2006]. Performance of methods for measurement of exposure to inorganic acids in workplace air. *J Environ Monit* 8:120-126.
- [15] Breuer D, Heckmann P, Gusbeth K, Schwab G, Blaskowitz M, Moritz A [2012]. Sulfuric acid at workplaces – applicability of the new indicative occupational exposure limit value (IOELV) to thoracic particles. *J Environ Monit* 14:440-445.
- [16] Comité Européen de Normalisation (CEN) [2012]. EN 482, Workplace exposure — General requirements for the performance of procedures for the measurement of chemical agents. Brussels, Belgium: Comité Européen de Normalisation (CEN).

METHOD WRITTEN BY:

Dietmar Breuer, Dr. rer. nat., Institute for Occupational Safety and Health of the German Social Accident Insurances, Sankt Augustin, Germany; and Kevin Ashley, Ph.D., NIOSH (retired)

METHOD REVISED BY:

Ronnee Andrews, Ph.D., Jennifer L. Roberts, and Peter Shaw, Ph.D., NIOSH

Disclaimer: Mention of any company or product does not constitute endorsement by the National Institute for Occupational Safety and Health, Centers for Disease Control and Prevention. In addition, citations to websites external to NIOSH do not constitute NIOSH endorsement of the sponsoring organizations or their programs or products. Furthermore, NIOSH is not responsible for the content of these websites. All web addresses referenced in this document were accessible as of publication date.