

Frequency Response of a Low-Power Closed-Path CO₂ and H₂O Eddy Covariance System

Steve Sargent and Hongcheol Kim, Campbell Scientific, Inc. Logan, Utah, USA

Abstract

Accurate measurement of CO₂ and H₂O eddy covariance fluxes requires an analyzer with good frequency response. Closed-path analyzers offer improved rain performance and the option of automatic zero and span, but their frequency response is generally worse than open-path analyzers due to the residence time in the sample cell and mixing along intake tubes and in the sample cell. H₂O frequency response can suffer additional losses due to interactions with the surface of the intake tube and sample cell. The traditional approach to maintain good frequency response in a closed-path system is to use a high sample flow rate and/or reduced pressure in the sample cell. This approach requires a relatively high power pump that makes it difficult to operate without access to AC mains power.

A new, small, closed-path CO₂ and H₂O Eddy Covariance System (CPEC200, Campbell Scientific, Inc.) achieves good frequency response (5.8 Hz half-power

er bandwidth) at low power (12 W total). The analyzer resides close to the sonic path to keep the intake tube short (0.6 m) and reduce mixing in the intake tube. The sample cell volume is small (5.8 ml) to minimize residence time (50 ms at 7 LPM sample flow). The entire system is designed for low pressure drop to reduce pump power.

The frequency response of the CPEC200 was measured by injecting an impulse of CO₂/H₂O at the system inlet. This measured impulse response was then Fourier transformed and normalized to give the frequency response. This frequency response was compared to a simple model that included a term representing ideal plug flow through the sample cell and an exponential mixing term.

Instrumentation Setup

Gas Analyzer: EC155 CO₂/H₂O Infrared Gas Analyzer

Intake Tube: Stainless steel, 0.105" ID x 23" (similar to standard heated intake tube)

Pump: CPEC200 Pump Module (pump speed controlled at 7 LPM)

Datalogging: CR3000 Micrologger

Measurement Rate: 50Hz from EC155 to CR3000 via SDM

Gas Injection: 1.5% CO₂, with fast injection valve

Sample Cell Pressure: 84.1 kPa (3.3 kPa below ambient pressure)



Figure 1. EC155 CO₂/H₂O Infrared Gas Analyzer.

Results

Time series of injected CO₂ and H₂O pulses are in remarkable agreement with each other indicating that the H₂O suffered little frequency response loss. This can be attributed to small sample cell volume (5.8 ml) and short intake tube (0.6 m) implemented in CPEC200 closed-path eddy covariance system with EC155 closed-path IRGA.




Figure 2. Normalized time series pulses. H₂O pulse has been inverted before overlaying with CO₂ pulse.

The graph shows good frequency response by both CO₂ and H₂O (5.8 Hz half-power bandwidth) which is excellent for low power used for the whole system (12 W total and 4 W for pump)




Figure 3. Overlay of CO₂ and H₂O spectra.

The CO₂ spectrum shows a good agreement with empirical model with a mixing term.




Figure 4. Display of CO₂ spectrum and empirical model with a mixing term.

Summary and Conclusions

- Excellent frequency responses were obtained for CO₂ and H₂O (5.8 Hz half-power bandwidth)
- Traditionally, frequency response of H₂O suffers loss due to interactions with the surface of the intake tube and sample cell. However, the frequency responses of CO₂ and H₂O show remarkable agreement with each other, indicating little loss in H₂O frequency response.
- Small sample cell volume and short intake tube preserve the frequency response of H₂O.
- The results show good agreement between the physical sample cell residence time (50 ms) and the modeled residence time (52 ms). The modeled exponential mixing time constant (18 ms) was much smaller than the residence time, indicating little mixing in the intake tube or the sample cell. H₂O frequency response was as good as that of CO₂ showing no difference between them.