Mauna Loa Solar Observatory Observer's Log ______ Fri Apr 19 16:48:39 GMT 2002 Year: 02 Doy: 109 Observer: yasukawa WEATHER COMMENT: Fri Apr 19 16:48:40 GMT 2002 cool, clear, light southwest breeze. Fri Apr 19 16:50:17 GMT 2002 CHIP Start Patrol Fri Apr 19 16:50:37 GMT 2002 PICS Start Patrol Fri Apr 19 16:50:56 GMT 2002 MKIV Start Patrol **PICS PROBLEM**: Fri Apr 19 16:57:49 GMT 2002 PICS objective wheel still flaky. Fri Apr 19 17:07:18 GMT 2002 PICS End Patrol PSPT COMMENT: Fri Apr 19 17:26:47 GMT 2002 Starting observations. **ECHO PROBLEM**: Fri Apr 19 17:27:17 GMT 2002 Image down in lower right corner. **ECHO PROBLEM**: Fri Apr 19 17:28:59 GMT 2002 Restarting echosys. HA error was -0.001 and DEC error was -0.000**PICS PROBLEM**: Fri Apr 19 17:47:55 GMT 2002 Checked out status board again and checked out home switch. Both appear OK. It looks like the failed component might be the HEI optical encoder. Swapping it out. Fri Apr 19 18:35:37 GMT 2002 PICS Start Patrol **PICS PROBLEM**: Fri Apr 19 18:35:54 GMT 2002 HEI optical encoder replaced with a new unit. Tests with handheld controller indicate problem is fixed. **ECHO PROBLEM**: Fri Apr 19 18:36:57 GMT 2002 Image is still offset in window. Restarting again. A NEW TAPE HAS BEEN PUT INTO KAIEE DLT DRIVE, Fri Apr 19 18:45:44 GMT 2002 Fri Apr 19 19:01:02 GMT 2002 CHIP LSD Fri Apr 19 19:03:18 GMT 2002 CHIP End LSD Fri Apr 19 19:03:32 GMT 2002 CHIP BiasLSD Fri Apr 19 19:04:35 GMT 2002 CHIP End BiasLSD Fri Apr 19 19:04:49 GMT 2002 CHIP Bias Fri Apr 19 19:05:48 GMT 2002 CHIP End Bias WEATHER COMMENT: Fri Apr 19 19:10:49 GMT 2002 In orographic clouds. PSPT COMMENT: Fri Apr 19 19:12:03 GMT 2002 Stopped observations. WEATHER COMMENT: Fri Apr 19 19:22:23 GMT 2002 Solid overcast. Closing dome in case of precipitation. Fri Apr 19 19:23:13 GMT 2002 CHIP End Patrol **MKIV PROBLEM**: Fri Apr 19 19:23:09 GMT 2002 Did not calibrate mk4 due to flurry of other activity so far this morning.

Fri Apr 19 19:24:21 GMT 2002 PICS End Patrol Fri Apr 19 19:24:40 GMT 2002 MKIV End Patrol

```
MKIV COMMENT: Fri Apr 19 21:35:24 GMT 2002
Connected the RTD and heater strip to the controller for FLC heater. Secured
cables inside the troughs. Secured the cable run between the wrap-up
and the AC switchbox via the service loop. Added the power cable to switch
box. Reinstalled the printing temperature monitor and its thermistor.
Buttoned up mk4. Waiting for 72-hour wet-out period to end before
firing up the new FLC heater system, should be Saturday 4/20.
WEATHER COMMENT: Fri Apr 19 21:43:16 GMT 2002
Still in orographic cloud overcast and light fog.
**ECHO PROBLEM**: Fri Apr 19 21:43:59 GMT 2002
Did not get a chance to see if image position got fixed by restart, clouds
too thick.
Fri Apr 19 21:46:37 GMT 2002
       MkIV
16 50.rawmk4 17 23.rawmk4 17 55.rawmk4 18 28.rawmk4 19 00.rawmk4
16 53.rawmk4 17 26.rawmk4 17 58.rawmk4 18 31.rawmk4 19 03.rawmk4
16 56.rawmk4 17 29.rawmk4 18 01.rawmk4 18 34.rawmk4 19 06.rawmk4
16 59.rawmk4 17 32.rawmk4 18 04.rawmk4 18 37.rawmk4 19 09.rawmk4
17_02.rawmk4 17_35.rawmk4 18_07.rawmk4 18_40.rawmk4 19_12.rawmk4
17_05.rawmk4 17_38.rawmk4 18_10.rawmk4 18_43.rawmk4 19_15.rawmk4
17 08.rawmk4 17 41.rawmk4 18 13.rawmk4 18 46.rawmk4 19 18.rawmk4
17_11.rawmk4 17_44.rawmk4 18_16.rawmk4 18_48.rawmk4 19_21.rawmk4
17 14.rawmk4 17 47.rawmk4 18 19.rawmk4 18 51.rawmk4
17 17.rawmk4 17 50.rawmk4 18 22.rawmk4 18 54.rawmk4
17_20.rawmk4 17_52.rawmk4 18_25.rawmk4 18_57.rawmk4
```