
Mauna Loa Solar Observatory Observer's Log

Mon Mar 26 15:35:02 GMT 2012

Year: 12 Doy: 086

Observer: berkey

Mon Mar 26 16:53:50 GMT 2012 H-ALPHA Start Patrol

WEATHER COMMENT: Mon Mar 26 16:53:55 GMT 2012

temp 32F, wind 7mph, sky looks mostly clear with light haze in the east

____end____

Mon Mar 26 16:55:26 GMT 2012 CHIP Start Patrol

GENERAL PROBLEM COMMENT BY BEN: Mon Mar 26 16:58:47 GMT 2012

nahenahe got a REPEATED LOGIN FAILURE error on the console.

____end____

Mon Mar 26 17:01:35 GMT 2012 COMP Start Patrol

Mon Mar 26 17:15:05 GMT 2012 MKIV Start Patrol

Mon Mar 26 17:25:57 GMT 2012: PSPT Start Patrol

Mon Mar 26 18:01:39 GMT 2012 CHIP LSD

Mon Mar 26 18:03:25 GMT 2012 CHIP End LSD

Mon Mar 26 18:03:35 GMT 2012 CHIP BiasLSD

Mon Mar 26 18:04:24 GMT 2012 CHIP End BiasLSD

Mon Mar 26 18:04:33 GMT 2012 CHIP Bias

Mon Mar 26 18:05:17 GMT 2012 CHIP End Bias

Mon Mar 26 18:05:24 GMT 2012 CHIP ReStart Patrol

Mon Mar 26 18:13:15 GMT 2012 COMP End Patrol

NOTE BY ALLEN: Mon Mar 26 20:14:33 GMT 2012

Reconfigured dome shutter.

____end____

Mon Mar 26 20:18:00 GMT 2012 MKIV End Patrol

Mon Mar 26 20:18:05 GMT 2012 MKIV Start Cal

CoMP COMMENT BY BEN: Mon Mar 26 20:23:15 GMT 2012

Comp observing is currently paused, to work in the implementation of a automatic occulter centering routine.

____end____

WEATHER COMMENT: Mon Mar 26 20:27:36 GMT 2012

Rapidly building orographic clouds are degrading images.

____end____

Mon Mar 26 20:30:27 GMT 2012: PSPT Abort Patrol

Mon Mar 26 20:31:15 GMT 2012 CHIP End Patrol

Mon Mar 26 20:33:10 GMT 2012 H-ALPHA End Patrol

Mon Mar 26 20:36:49 GMT 2012 MKIV End Cal

Mon Mar 26 20:37:24 GMT 2012 MKIV End Patrol

WEATHER COMMENT: Mon Mar 26 20:37:35 GMT 2012

In clouds, instruments idled.

____end____

NOTE BY BEN: Mon Mar 26 21:11:36 GMT 2012

Per "Matrix-UPS Displays Overload on Display Panel" document on the apc webpage

The apc display was reset by unscrewing the large "thumb" screw in the back of the unit between the electroics and isolati
on units. Once the relay clicked off the thumb screw was tightened and the relay clicked back on. Display became blank

until button was pushed and now it looks "happy" again.

____end____

COMP COMMENT BY BEN: Tue Mar 27 01:12:16 GMT 2012

COMP vi has been reverted to pervious know working code, so it should be ready for tomorrow

____end____

NOTE BY BEN: Tue Mar 27 01:45:59 GMT 2012

Fixed display on APC ups by resetting the electronics unit.

____end____

Tue Mar 27 01:50:53 GMT 2012

MkIV

17_15.rawmk4	17_56.rawmk4	18_37.rawmk4	19_18.rawmk4	19_59.rawmk4
17_18.rawmk4	17_59.rawmk4	18_40.rawmk4	19_21.rawmk4	20_02.rawmk4
17_21.rawmk4	18_02.rawmk4	18_43.rawmk4	19_24.rawmk4	20_05.rawmk4
17_24.rawmk4	18_05.rawmk4	18_46.rawmk4	19_27.rawmk4	20_08.rawmk4
17_26.rawmk4	18_08.rawmk4	18_49.rawmk4	19_30.rawmk4	20_11.rawmk4
17_29.rawmk4	18_11.rawmk4	18_52.rawmk4	19_33.rawmk4	20_14.rawmk4
17_32.rawmk4	18_14.rawmk4	18_55.rawmk4	19_36.rawmk4	20_21.rawmk4
17_35.rawmk4	18_17.rawmk4	18_58.rawmk4	19_39.rawmk4	20_27.rawmk4
17_38.rawmk4	18_19.rawmk4	19_01.rawmk4	19_42.rawmk4	20_33.rawmk4
17_41.rawmk4	18_22.rawmk4	19_04.rawmk4	19_45.rawmk4	c20_18.rawmk4
17_44.rawmk4	18_25.rawmk4	19_07.rawmk4	19_48.rawmk4	c20_24.rawmk4
17_47.rawmk4	18_28.rawmk4	19_09.rawmk4	19_51.rawmk4	c20_30.rawmk4
17_50.rawmk4	18_31.rawmk4	19_12.rawmk4	19_53.rawmk4	
17_53.rawmk4	18_34.rawmk4	19_15.rawmk4	19_57.rawmk4	