
Mauna Loa Solar Observatory Observer's Log

Sun Feb 2 17:13:57 GMT 2014

Year: 14 Doy: 033

Observer: waters

WEATHER COMMENT: waters: Sun Feb 2 17:14:06 GMT 2014

Vog is at about 9,000ft. Clear skies with low clouds coming in from the West, Northwest. Temp is 39F with windspeed at 16m
ph from the South, South-west. Still snow on the ground.

____end____

Sun Feb 2 17:43:05 GMT 2014: PSPT Start Patrol

Sun Feb 02 18:06:49 GMT 2014 COMP Start Patrol

Sun Feb 02 18:12:22 GMT 2014 KCOR Start Synoptic Patrol

Sun Feb 02 19:04:59 GMT 2014 KCOR End Patrol

Sun Feb 02 19:05:00 GMT 2014 KCOR Start Calibration script: c:\kcor\mlso-calibration22deg.ini

Sun Feb 02 19:20:48 GMT 2014 KCOR End Calibration Script

Sun Feb 02 19:21:38 GMT 2014 KCOR Start Synoptic Patrol

Sun Feb 02 19:25:39 GMT 2014 COMP End Patrol

Sun Feb 02 19:25:40 GMT 2014 COMP Start Patrol

Sun Feb 02 20:43:15 GMT 2014 COMP End Patrol

Sun Feb 02 20:43:16 GMT 2014 COMP Start Patrol

Sun Feb 02 20:45:14 GMT 2014 COMP End Patrol

Sun Feb 02 20:47:39 GMT 2014 COMP Start Patrol

Sun Feb 02 20:48:32 GMT 2014 COMP End Patrol

Sun Feb 02 20:49:28 GMT 2014 COMP Start Patrol

GENERAL OBSERVATORY COMMENT BY waters: Sun Feb 2 20:49:42 GMT 2014

Sgs adjusted.

____end____

Sun Feb 02 22:07:02 GMT 2014 COMP End Patrol

Sun Feb 02 22:07:04 GMT 2014 COMP Start Patrol

Sun Feb 02 22:39:07 GMT 2014 COMP End Patrol

Sun Feb 02 22:40:44 GMT 2014 COMP Start Patrol

GENERAL OBSERVATORY COMMENT BY waters: Sun Feb 2 22:41:02 GMT 2014

Sgs adjusted.

____end____

Sun Feb 02 22:54:06 GMT 2014 KCOR End Patrol

Sun Feb 02 22:54:20 GMT 2014 COMP End Patrol

GENERAL OBSERVATORY COMMENT BY waters: Sun Feb 2 22:54:20 GMT 2014

Clouds have rolled in. Comp, shabar and kcor ilded.

____end____

Sun Feb 2 23:01:22 GMT 2014: PSPT Abort Patrol

GENERAL OBSERVATORY COMMENT BY waters: Sun Feb 2 23:02:36 GMT 2014

Fog has rolled in, Pspt ilded and domes closed.

____end____

GONG COMMENT BY waters: Sun Feb 2 23:22:33 GMT 2014

Went and inspected Gong. Flipped white circuit breaker and the bottom of the rack and restarted Data and instrument racks. Iso drive turned on without a hitch but when the Data rack was turned on, the iso drive shut off again. Not too sure what is going on. Tried the circuit breaker and restart again but to no avail.

__end__