
Mauna Loa Solar Observatory Observer's Log

Tue Feb 23 17:27:11 GMT 2016

Year: 16 Doy: 054

Observer: stueben

WEATHER COMMENT: stueben: Tue Feb 23 17:28:56 GMT 2016

Thin cirrus to the east, clearer west, wind south 10-15 mph, temp 49F.

____end____

KCOR COMMENT BY stueben: Tue Feb 23 18:06:50 GMT 2016

Kcor is running but images are poor due to thin cirrus.

____end____

Tue Feb 23 18:43:58 GMT 2016 KCOR Start Synoptic Patrol

WEATHER COMMENT: stueben: Tue Feb 23 18:45:19 GMT 2016

Sky is still a bit bright, but cirrus has thinned enough for decent kcor images.

____end____

Tue Feb 23 19:29:24 GMT 2016 SGS Alignment complete

Tue Feb 23 20:49:28 GMT 2016 KCOR End Patrol

Tue Feb 23 20:49:29 GMT 2016 KCOR Start Calibration script: c:\kcor\mlso-calibration22deg-20150323.ini

Tue Feb 23 21:06:31 GMT 2016 KCOR End Calibration Script

Tue Feb 23 21:06:48 GMT 2016 KCOR Start Synoptic Patrol

Tue Feb 23 21:06:48 GMT 2016 KCOR Start Synoptic Patrol

Tue Feb 23 22:48:45 GMT 2016 SGS Alignment complete

Tue Feb 23 22:47:12 GMT 2016 KCOR End Patrol

Tue Feb 23 23:10:42 GMT 2016 KCOR Start Synoptic Patrol

GENERAL COMMENT BY stueben: Tue Feb 23 23:14:05 GMT 2016

Lost power to sgs for a few minutes when one of the workers inadvertently shut off a circuit breaker when removing the cover of the breaker box to put up the new panelling.

____end____

Wed Feb 24 00:52:43 GMT 2016 KCOR Start Synoptic Patrol

GENERAL PROBLEM COMMENT BY stueben : Wed Feb 24 01:01:52 GMT 2016

Another circuit breaker inadvertently tripped - this time hale and mlso server went down. This shouldn't happen as they are on the big APC ups, but the ups seems to be stuck in the mode where it just passes through the line power. The ups display is not working so its difficult to determine exactly what it is doing.

____end____

Wed Feb 24 02:10:16 GMT 2016 KCOR End Patrol

GENERAL COMMENT BY stueben: Wed Feb 24 02:19:27 GMT 2016

The contractor worked on the walls and ceilings on the north side of the control room today.

____end____