


May 27, 2016



6312 NW 18th Drive
Suite 100
Gainesville, FL 32653

352.378.0332 PHONE
352.378.0354 FAX

www.TRCSolutions.com

May 27, 2016

Ms. Lori Simmons
Arkansas Department of Health
4815 West Markham Street
Little Rock, Arkansas 72205
Via email Lori.Simmons@arkansas.gov

Re: Georgia-Pacific, Crossett Mill - Biweekly Air Monitoring Report for Hydrogen Sulfide

Dear Ms. Simmons,

Following is the biweekly data summary for the Georgia-Pacific (GP) hydrogen sulfide (H₂S) and meteorological monitoring program, at the GP Crossett mill, covering the calendar period of May 3rd through May 17th.

Summary of Results

Included in this report are three plots presenting H₂S concentrations calculated with varied rolling average periods (30-minute, 8-hour, and 24-hour).

Also included in this report is a summary of results from the daily 1-point QC checks performed during this biweekly period. The QAPP establishes goals for precision and bias as a coefficient of variation (CV) <10% and ± 10%, respectively. Precision and bias are calculated in accordance with 40 CFR Part 58 Appendix A, Section 4.1.

There was a single occurrence of data loss during this two week period, other than those resulting from automated daily 1-point QC and weekly calibration checks. On the morning of the 14th, the internal logger on the analyzer malfunctioned. The logger malfunction resulted in an extended period of data loss (approximately 38 hours); the analyzer was reset on April 15th. Due to the logger malfunction, automated calibration checks were not recorded on the 14th and 15th. Results for all available automated daily 1-point QC checks fall within the acceptable range, indicating the H₂S monitor was operating in accordance with the QAPP.

Fourteen-day time series plots for all recorded meteorological (met) parameters are presented in the final table. All met parameters have 100% data capture for this report period.



May 27, 2016

Please feel free to contact me if you have any questions or need any additional data.

Sincerely,




Jonathan Bowser
Manager, Air Quality and Meteorological Monitoring


Air Measurements – Gainesville Office
6312 NW 18th Drive, Suite 100
Gainesville, Florida 32653
(352) 260-1162
Email: jbowser@trcsolutions.com

CC: Becky Keough, ADEQ Director via email: keogh@adeq.state.ar.us
Kara Allen, Environmental Engineer, USEPA Region 6 via email Allen.Kara@epa.gov


H₂S 30 Min Rolling Avg Georgia Pacific Crossett, AR



H₂S 8 Hr Rolling Avg Georgia Pacific Crossett, AR




H₂S 24 Hr Rolling Avg Georgia Pacific Crossett, AR



H₂S Assessment

GP - Crossett, AR			Constituent type: H ₂ S					CV _{ub} (%)	Bias (%)
Date	Meas Val (Y)	Audit Val (X)	d (Eqn. 1)	25th Percentile	d ²	d	d ²		
5/3/2016 13:00	70.1	70.0	0.1	-0.464	0.020	0.143	0.020		
5/4/2016 13:00	69.8	70.0	-0.3	75th Percentile	0.082	0.286	0.082		
5/5/2016 13:00	69.7	70.0	-0.4	0.036	0.184	0.429	0.184	n	S_d
5/6/2016 13:00	69.8	70.0	-0.3		0.082	0.286	0.082	12	0.301
5/7/2016 13:00	69.6	70.0	-0.6		0.327	0.571	0.327	n-1	Σd
5/8/2016 13:00	69.9	70.0	-0.1		0.020	0.143	0.020	11	-2.714
5/9/2016 13:00	70.1	70.0	0.1		0.020	0.143	0.020		S_{d2}
5/10/2016 13:00	70.1	70.0	0.1		0.020	0.143	0.020		0.167
5/11/2016 13:00	69.9	70.0	-0.1		0.020	0.143	0.020		Σ d
5/12/2016 13:00	70.0	70.0	0.0		0.000	0.000	0.000		3.571
5/13/2016 13:00	69.6	70.0	-0.6		0.327	0.571	0.327		"AB" (Eqn 4)
5/16/2016 13:00	69.5	70.0	-0.7		0.510	0.714	0.510		0.298
									"AS" (Eqn 5)
									0.223
									Bias (%) (Eqn 3)
									0.41
									Both Signs Positive
									FALSE
									Signed Bias (%)
									+/-0.41
									Both Signs Negative
									FALSE
									CV (%) (Eqn 2)
									0.42
									Upper Probability Limit
									0.36
									Lower Probability Limit
									-0.82



Meteorological Summary

