


January 10, 2018



6312 NW 18th Drive
Suite 100
Gainesville, FL 32653

352.378.0332 PHONE
352.378.0354 FAX

www.TRCSolutions.com

January 10, 2018

Ms. Lori Simmons
Arkansas Department of Health
4815 West Markham Street
Little Rock, Arkansas 72205
Via email Lori.Simmons@arkansas.gov

Re: Georgia-Pacific, Crossett Mill - Biweekly Air Monitoring Report for Hydrogen Sulfide

Dear Ms. Simmons,

Please find the following biweekly report for the Georgia-Pacific (GP) Crossett Mill hydrogen sulfide (H₂S) and meteorological monitoring program covering the calendar period of December 13, 2017 through December 26, 2017.

Summary of Results

Included in this report are three plots presenting H₂S concentrations across different rolling average periods (30-minute, 8-hour, and 24-hour), daily 1-point quality control (QC) checks with precision and bias estimates and time series plots for all recorded meteorological (met) parameters for the two week period.

Data Quality

The Quality Assurance Project Plan (QAPP) establishes measurement quality objectives (MQOs) for H₂S regarding precision and bias expressed as a coefficient of variation (CV) <10% and $\pm 10\%$, respectively. Precision and bias are calculated in accordance with 40 CFR Part 58 Appendix A, Section 4.1. Precision and bias calculations are presented on page six of this report.

Results for available automated daily 1-point QC checks were within the accuracy objective, $\pm 10\%$, indicating the H₂S monitor was operating in accordance with MQOs as stated in the QAPP.

Additionally, weekly automated zero adjustments were implemented starting February 1, 2017. During this reporting period two automated zero checks were performed; within the acceptable range



of ± 1.5 ppb, as defined in the QAPP. The result for these zero checks are presented below.

Date	Zero Check Response (ppb)
12/14/2017	0.1
12/21/2017	0.6

On December 26th, TRC personnel observed negative drift while reviewing incoming data. TRC remotely accessed the H₂S analyzer to perform a zero adjustment followed by a three point calibration check.

Data Capture

There were multiple occurrences of H₂S data loss this monitoring period, in addition to those resulting from automated daily 1-point QC and weekly calibration checks. A server error on December 24th was responsible for the loss of approximately four hours of data in the morning. On December 26th, TRC personnel performed a remote zero adjustment and calibration check resulting in approximately four hours of invalid H₂S data. The activity on the 26th interrupted the scheduled automated calibration check on that day.

Fourteen-day time series plots for all recorded meteorological (met) parameters are presented in the final table. There was a single occurrence of met data loss during this monitoring period. On December 24th all met parameters were lost for nearly five and a half hours on account of a server error.

Please feel free to contact me if you have any questions or need any additional data.

Sincerely,




Jonathan Bowser
Manager, Air Quality and Meteorological Monitoring


Air Measurements – Gainesville Office
6312 NW 18th Drive, Suite 100
Gainesville, Florida 32653
(352) 260-1162
Email: jbowser@trcsolutions.com

CC: Becky Keough, ADEQ Director via email: keogh@adeq.state.ar.us
Kara Allen, Environmental Engineer, USEPA Region 6 via email Allen.Kara@epa.gov


H2S 30 Min Rolling Avg Georgia Pacific Crossett, AR



H2S 8 Hr Rolling Avg Georgia Pacific Crossett, AR



H2S 24 Hr Rolling Avg Georgia Pacific Crossett, AR




H₂S Assessment

GP - Crossett, AR			Compound of Interest: H ₂ S				CV _{ub} (%)	Bias (%)																				
Date	Meas Val (Y)	Input Val (X)	d (Eqn. 1)	25th Percentile	d ²	d	d ²																					
12/13/2017 13:00	66.7	70.0	-4.7	-4.143	22.224	4.714	22.224																					
12/14/2017 13:00	67.1	70.0	-4.1	75th Percentile	17.163	4.143	17.163	<table border="1"> <tr> <td>n</td> <td>S_d</td> <td>S_{d2}</td> <td>Σ d </td> <td>"AB" (Eqn 4)</td> </tr> <tr> <td>13</td> <td>1.661</td> <td>12.134</td> <td>42.143</td> <td>3.242</td> </tr> <tr> <td>n-1</td> <td>Σd</td> <td>Σd²</td> <td>Σ d ²</td> <td>"AS" (Eqn 5)</td> </tr> <tr> <td>12</td> <td>-42.143</td> <td>169.735</td> <td>169.735</td> <td>1.661</td> </tr> </table>	n	S _d	S _{d2}	Σ d	"AB" (Eqn 4)	13	1.661	12.134	42.143	3.242	n-1	Σd	Σd ²	Σ d ²	"AS" (Eqn 5)	12	-42.143	169.735	169.735	1.661
n	S _d	S _{d2}	Σ d	"AB" (Eqn 4)																								
13	1.661	12.134	42.143	3.242																								
n-1	Σd	Σd ²	Σ d ²	"AS" (Eqn 5)																								
12	-42.143	169.735	169.735	1.661																								
12/15/2017 13:00	66.5	70.0	-5.0	-2.286	25.000	5.000	25.000																					
12/16/2017 13:00	67.1	70.0	-4.1		17.163	4.143	17.163																					
12/17/2017 13:00	68.2	70.0	-2.6		6.612	2.571	6.612																					
12/18/2017 13:00	68.4	70.0	-2.3		5.224	2.286	5.224																					
12/19/2017 13:00	69.1	70.0	-1.3		1.653	1.286	1.653																					
12/20/2017 13:00	68.8	70.0	-1.7		2.939	1.714	2.939																					
12/21/2017 13:00	68.4	70.0	-2.3		5.224	2.286	5.224																					
12/22/2017 13:00	69.4	70.0	-0.9		0.735	0.857	0.735																					
12/23/2017 13:00	68.2	70.0	-2.6		6.612	2.571	6.612																					
12/24/2017 13:00	67.2	70.0	-4.0		16.000	4.000	16.000																					
12/25/2017 13:00	65.4	70.0	-6.6		43.184	6.571	43.184																					

Bias (%) (Eqn 3)	Both Signs Positive
4.06	FALSE
Signed Bias (%)	Both Signs Negative
-4.06	TRUE

CV (%) (Eqn 2)	2.29
-----------------------	------

Upper Probability Limit	Lower Probability Limit
0.01	-6.5



Meteorological Summary

