




Vacuum Trap Assembly Quick Guide



- Aspirator is connected to the collection flask (A) which is connected to the overflow flask (B)
- Both A & B must contain appropriate disinfectant (10-20% bleach solution); In Addition, an anti-foaming solution should be considered for the overflow flask (B)
- Both A & B are placed inside secondary containment(s)
- HEPA filter (C) is placed between the overflow flask (B) and the vacuum source (D) to prevent contamination of the vacuum line

Questions?
Email safety@andrew.cmu.edu