

The World Infrared Standard Group

Status report




Julian Gröbner, Christian Thomann

Physikalisch-Meteorologisches Observatorium Davos und
World Radiation Center (PMOD/WRC),
Davos Dorf, Switzerland

The Earth global radiation balance


Why we need SI traceable radiation measurements




Wild et al., 2015

These global energy balance estimates are part of the newest IPCC AR6 report, published in August 2021.

Pyrgeometer energy balance



Eppley PIR



Kipp&Zonen CG4

Typical values for clear nights:

Net outgoing radiation $\sim 100 \text{ Wm}^{-2}$

PIR Dome ($\Delta T=1\text{K}$) $\sim 18 \text{ Wm}^{-2}$

Pyrgeometer equations

Several choices for the radiometric model of a pyrgeometer:

Simple Albrecht formula

$$E = \frac{U}{C} + \sigma T_{BODY}^4 - K \sigma (T_{DOME}^4 - T_{BODY}^4)$$

Extended Albrecht
(PMOD)

$$E = \frac{U}{C} (1 + k_1 \sigma T_{BODY}^3) + k_2 \sigma T_{BODY}^4 - k_3 \sigma (T_{DOME}^4 - T_{BODY}^4)$$

NREL formula

$$E = K_0 + K_1 U + K_2 \sigma T_{RECEIVER}^4 - K_3 (T_{DOME}^4 - T_{RECEIVER}^4)$$

Fit parameters




Figure A1: This is a partial copy of Fig 1 from Philipona et al (1995). Note that it does not show the reflection from the sensor surface.


Pyrgeometer calibration procedure

Described in IOM 120 (Gröbner and Wacker, 2015)

$$E = \frac{U}{C} \left(1 + k_1 \sigma T_B^3 \right) + k_2 \sigma T_B^4 - k_3 \sigma (T_D^4 - T_B^4)$$

- 1) Determination of k_1 , k_2 , k_3 in Blackbody (instrument constants)
- 2) Outdoor measurements relative to WISG to retrieve responsivity C


$$C = \frac{U(1 + k_1 \sigma T_B^3)}{k_2 \sigma T_B^4 - k_3 \sigma (T_D^4 - T_B^4) - E_{WISG}}$$



Calibration criteria

1. Outliers are removed ($U > 0.001$ V, $U < -20$ mV, $|T_D| > 40$ °C, $|T_B| > 40$ °C)
2. Any night containing rain is excluded (limit of 0.2 mm/10 min)
3. Stable atmospheric conditions, defined by the standard deviation of the WISG < 2 Wm $^{-2}$
4. Net radiation measured by the WISG < -70 Wm $^{-2}$
5. Measurements from one night are used if there are at least 80% valid measurement points
6. Night is defined when the solar zenith angle is larger than 95°
7. Relative standard deviation of the test pyrgeometer signal $< 3\%$
8. Integrated water vapour (IWV) greater than 10 mm

Characterisation in the blackbody to retrieve instrument constants k_1 , k_2 , and k_3



Pyrgeometer temperature range: +20°C to -10°C

Black Body temperature range: +15°C to -20°C

```
bbcalib('31464f3_18122016.dat');
```

Albrecht Formula:

$C=3.651$ or using SVD $C=3.6510+-0.0067$

$K(k3)=2.819$ or using SVD $K=2.819+-0.0381$

Uncertainty of C for $k1$ (+-0.012000): 0.006000

Uncertainty of C for $k2$ (+-0.000400): 0.012126

Uncertainty of C for $k3$ (+-0.015000): 0.003694

Uncertainty of C for U (+-0.000000): 0.001059

Uncertainty of C for E (+-0.090000): 0.007747

Uncertainty of C for Tb (+-0.010000): 0.015791

Uncertainty of C for Td (+-0.010000): 0.011548

Total uncertainty: 0.0253

SVD: $C=3.9154+-0.0494$


SVD: $k1=0.03291+-0.01338$

SVD: $k2=0.99544+-0.00038$

SVD: $k3=2.72064+-0.01000$

Expanded Uncertainty ($k=2$) of C (microV) : 0.0531


Results from the regression $\rightarrow C_{BB}$, k_1, k_2, k_3



Blackbody temperatures min=-19.3, max=14.9 degC


PIR Body temperatures min=-9.3, max=20.4 degC

Ambient temperatures min=22.9, max=23.8 degC



Outdoor calibration relative to the WISG to retrieve the responsivity C




$$C = \frac{U(1+k_1\sigma T_B^3)}{k_2\sigma T_B^4 - k_3\sigma(T_D^4 - T_B^4) - E_{WISG}}$$



$$C=12.319 \pm 0.20 \mu\text{VW}^{-1}\text{m}^2$$


$$u=1.6\%$$

Standard uncertainty relative to WISG 0.3%




Pyrgeometers calibrated in the blackbody give inconsistent results when measuring atmospheric longwave irradiance

Responsivity ratio from Blackbody versus WISG




From: IPgC-II report, 2017



Differences between BB & WISG based calibrations are linked to the spectral inhomogeneities of the pyrgeometers (dome transmission & thermopile absorptivity).

Gröbner and Los, 2007

The World Infrared Standard Group (WISG)



Consists of 4 Pyrgeometers:

- 2 modified Eppley PIR, s/n 31463, 31464
- 2 Kipp & Zonen CG4, s/n FT004, 010535 (since 1 June 2004)

The WISG has been stable to $\pm 1 \text{ W m}^{-2}$ since 2004

Traceability of atmospheric longwave irradiance to SI




BB2007



- Cylindrical cavity
- effective emissivity
 $0.99993(33)$

Gröbner, AO 2008

IRIS



- Windowless
- Flat Spectral Response
- Nighttime operation only

Gröbner, Metrologia, 2012

WISG



- PIR31463F3
- PIR31464F3
- CG4 FT004
- CG4 010535


Gröbner et al., JGR, 2014




WMO, IOM 120

Traceability of atmospheric longwave irradiance to SI

Comparison to the national temperature scale of PTB




Radiation thermometer



Comparison WISG to IRIS 2016-2020


Difference WISG to IRIS 4 in Wm^{-2} for IWV>10 mm

Instrument	Mean	STD	U95	N
WISG1 PIR 31463F3	-4.5	1.3	[-7.4 -1.6]	16495
WISG2 PIR 31464F3	-4.8	1.2	[-7.6 -2.2]	16495
WISG3 CG4 FT004	-4.8	1.2	[-7.5 -1.9]	15375
WISG4 CG4 010535	-4.8	1.2	[-7.4 -1.9]	15375
WISG avg	-4.7			



Confirms results from Gröbner et al., 2014 with IRIS & ACP


IPgC-III







- 37 Pyrgeometer
- 4 IRIS
- 4 ACP

- 3 Eppley PIR
- 24 OTT CG4/CGR4
- 6 Hukseflux IR20/IRxx
- 2 EKO MS-20/MS-21
- 2 EMPIR

IRIS & ACP




IRIS & ACP96 using Bruce Forgan analysis



Development of a spectrally flat pyrgeometer

Instrument features:

- Diamond dome
- Thermopile coated with Vertically Aligned Carbon Nanotubes (VACN)



IPgC-II report & calibration certificates

Instruments and Observing Methods
Report No. 129

Report on the Second International
Pyrgometer Intercomparison
(27 Sept - 15 Oct 2015, PMOD/WRC)

J. Gröbner and C. Thomann (Switzerland)



Calibration Certificate

No. 2015-112

Calibration Item

Pyrgometer

Manufacturer
Type

The Eppley Laboratory, Inc.
Precision Infrared Radiometer, with three dome thermistor

Serial number

Customer

Calibration Mark 2015-1129-01

Period of Calibration 29-Sep-2015 to 12-Oct-2015

Davos Dorf, 05 November, 2015

C. Thomann
In charge of calibration

Dr. Julian Gröbner
Head IR Radiometry Section


Calibration certificates without signature are not valid. This calibration certificate shall not be reproduced except in full without the written approval of the Physikalisch-Meteorologisches Observatorium Davos and World Radiation Center.

IPgC-II report & calibration certificates

Instruments and Observing Methods
Report No. 129

Report on the Second International
Pyrgeometer Intercomparison
(27 Sept - 15 Oct 2015, PMOD/WRC)

J. Gröbner and C. Thomann (Switzerland)



pmod wrc

Governance and traceability of atmospheric longwave irradiance

Annex to Resolution 1 (CIMO 17)

According to its Terms of Reference, in response to the requirement for standardization of atmospheric longwave radiation measurements, the Commission for Instruments and Methods of Observation (CIMO) decides to establish a governance framework for the World Infrared Standard Group (WISG).

The Governance framework comprises an advisory group of at least, five experts in atmospheric longwave radiation measurements, appointed by the president of CIMO for each International Pyrgeometer Comparison, preferably from among the participants in the comparison.

The comparison's leader, appointed by PMOD, will be invited to participate in the group's meeting.

The tasks of the advisory group are, but not limited to:

- (a) To review the status and stability of the WISG, and evaluate its role as operational reference standard for providing a stable longwave reference, based on the analysis provided by PMOD/WRC;
- (b) To recommend the updating of the calibration factors and changes to the WISG, if necessary;
- (c) To ensure the supervision of the International Pyrgeometer Comparison, scheduled to take place every five years in conjunction with the International Pyrheliometer Comparison;
- (d) To review progress in and provide advice on maintaining and improving traceability to the SI through the International Pyrgeometer Comparison;
- (e) To report their findings and recommendations to the CIMO Management Group.

WMO, No. 1227, 2018